

DATE

9/12/79

AIRCRAFT

C130

FLIGHT

790912 F

A. T01244

INSTRUMENT STATUS AND PERFORMANCE

	PreFlight	InFlight	PostFlight	Remarks	Data Units Collected
Johnson Williams	✓				
Nimbliometer	✓				
Lyman Alpha	N/A				
U. V.	✓				
dewpoint	✓				
Foil Impactor	✓				
Formvar	N/A				
Knollenberg					
Raindrop	N/A				
Cloud Droplet	✓				
FSSP	N/A				
Data System	✓				
& Displays	✓				
Ice Particle Counter					
Mee	✓				
ERT	N/A				
CO ₂ Radiometer	N/A				
Microwave Radiometer					
Aerosol					
Filters					
Bulk Water					
INC					
CCN					

B.

REMARKS

1.) foil started to run erratic (~600")

He stopped 681" at 1450Z

no data all T/O + ferry flight

2) tape 1 no EOF tape 2 is just continuation of tape 1, changed at ~1700Z

3) on Knollenberg only channels 1, 4, 5 ~~seen~~
feed out on monitor, don't know about
other channels (at ~1800 could monitor
all channels, must be working now?)

4) off 20Z

PAGE

TO 1244

FOIL MARKED AT SLOT AND COUNTER ZEROED.

OPERATOR

[illegible]