

# SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 2980  
CALIBRATION DATE: 14-Aug-07

SBE4 CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Seimens/meter

## GHIJ COEFFICIENTS

g = -1.00324949e+001  
h = 1.36939830e+000  
i = 5.42928292e-004  
j = 2.98567526e-005  
CPcor = -9.5700e-008 (nominal)  
CTcor = 3.2500e-006 (nominal)

## ABCDM COEFFICIENTS

a = 5.09561555e-004  
b = 1.36932362e+000  
c = -1.00318662e+001  
d = -8.26696850e-005  
m = 3.2  
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.70503	0.00000	0.00000
-1.0002	34.9478	2.81416	5.27269	2.81412	-0.00004
0.9998	34.9475	2.98609	5.38998	2.98612	0.00003
14.9998	34.9479	4.28602	6.20481	4.28608	0.00005
18.4998	34.9482	4.63395	6.40518	4.63391	-0.00004
28.9999	34.9441	5.72088	6.99419	5.72085	-0.00002
32.4999	34.9365	6.09456	7.18549	6.09458	0.00002

Conductivity =  $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$  Siemens/meter

Conductivity =  $(af^m + bf^2 + c + dt) / [10 (1 + \epsilon p)]$  Siemens/meter

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

