

# SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3338  
CALIBRATION DATE: 27-Feb-08

SBE4 CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Seimens/meter

## GHIJ COEFFICIENTS

g = -9.97710595e+000  
h = 1.54292595e+000  
i = -2.29418227e-003  
j = 2.69175391e-004  
CPcor = -9.5700e-008 (nominal)  
CTcor = 3.2500e-006 (nominal)

## ABCDM COEFFICIENTS

a = 1.13039470e-006  
b = 1.53702865e+000  
c = -9.96558662e+000  
d = -8.26445068e-005  
m = 6.0  
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.54629	0.00000	0.00000
-1.0000	34.9366	2.81336	4.97759	2.81334	-0.00002
1.0000	34.9367	2.98527	5.08857	2.98529	0.00002
15.0000	34.9377	4.28493	5.85940	4.28493	0.00000
18.5000	34.9373	4.63268	6.04890	4.63271	0.00002
29.0000	34.9359	5.71970	6.60578	5.71964	-0.00006
32.5000	34.9302	6.09360	6.78670	6.09364	0.00004

Conductivity =  $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$  Siemens/meter

Conductivity =  $(af^m + bf^2 + c + dt) / [10 (1 + \epsilon p)]$  Siemens/meter

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

