

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3338
CALIBRATION DATE: 16-Jul-08

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Seimens/meter

GHIJ COEFFICIENTS

g = -1.01945761e+001
h = 1.57722667e+000
i = -2.63636817e-003
j = 2.93538385e-004
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 5.94621971e-007
b = 1.57021544e+000
c = -1.01801738e+001
d = -7.72814079e-005
m = 6.3
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.54625	0.00000	0.00000
-1.0001	34.9346	2.81321	4.93869	2.81321	0.00000
0.9999	34.9352	2.98515	5.04821	2.98515	-0.00000
14.9999	34.9360	4.28473	5.80920	4.28472	-0.00001
18.4999	34.9357	4.63248	5.99636	4.63249	0.00001
28.9999	34.9345	5.71948	6.54653	5.71949	0.00000
32.4999	34.9302	6.09359	6.72534	6.09358	-0.00000


Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction



16-Jul-08 1.0000000