

# SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 2980  
CALIBRATION DATE: 12-Mar-10

SBE4 CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Seimens/meter

## GHIJ COEFFICIENTS

g = -1.00423934e+001  
h = 1.37226406e+000  
i = -1.39915596e-004  
j = 7.52228638e-005  
CPcor = -9.5700e-008 (nominal)  
CTcor = 3.2500e-006 (nominal)

## ABCDM COEFFICIENTS

a = 6.13880803e-005  
b = 1.37188623e+000  
c = -1.00416490e+001  
d = -8.36328941e-005  
m = 4.0  
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.70503	0.00000	0.00000
-1.0001	34.8128	2.80431	5.26560	2.80432	0.00000
0.9999	34.8132	2.97572	5.38264	2.97571	-0.00001
14.9999	34.8127	4.27121	6.19583	4.27120	-0.00000
18.4999	34.8121	4.61786	6.39582	4.61787	0.00001
28.9999	34.8107	5.70149	6.98386	5.70148	-0.00001
32.5000	34.8056	6.07433	7.17498	6.07434	0.00001

Conductivity =  $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$  Siemens/meter

Conductivity =  $(af^m + bf^2 + c + dt) / [10 (1 + \epsilon p)]$  Siemens/meter

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

