

SEA-BIRD ELECTRONICS, INC.
1808 136th Place N.E., Bellevue, Washington, 98005 USA
Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0140
CALIBRATION DATE: 27-Jul-05p

SBE 43 OXYGEN CALIBRATION DATA

COEFFICIENTS

Soc = 0.3102

Boc = 0.0000

Voffset = -0.6160

TCor = 0.0016

PCor = 1.350e-04

BATH OX (ml/l)	BATH TEMP ITS-90	BATH SAL PSU	INSTRUMENT OUTPUT(VOLTS)	INSTRUMENT OXYGEN(ml/l)	RESIDUAL (ml/l)
1.30	20.00	0.01	1.259	1.31	0.00
1.31	26.00	0.01	1.335	1.32	0.00
1.33	12.00	0.01	1.177	1.33	0.01
1.33	2.00	0.00	1.057	1.33	-0.00
1.33	30.00	0.01	1.393	1.33	0.00
1.33	6.00	0.00	1.107	1.34	0.00
4.13	20.00	0.01	2.650	4.13	0.01
4.15	26.00	0.01	2.886	4.15	0.00
4.17	12.00	0.01	2.373	4.18	0.01
4.18	30.00	0.01	3.046	4.17	-0.01
4.19	6.00	0.00	2.155	4.19	0.00
4.23	2.00	0.00	2.012	4.20	-0.03
6.84	6.00	0.00	3.129	6.85	0.01
6.84	12.02	0.01	3.502	6.87	0.03
6.90	2.00	0.00	2.901	6.87	-0.02
6.93	30.00	0.01	4.648	6.92	-0.02
6.97	26.00	0.01	4.422	6.96	-0.01
6.98	20.00	0.01	4.059	7.00	0.02


oxygen (ml/l) = (Soc * (V + Voffset)) * exp(Tcor * T) * Oxsat(T,S) * exp(PCor * P)

V = voltage output from SBE43, T = temperature [deg C], S = salinity [PSU]

Oxsat(T,S) = oxygen saturation [ml/l], P = pressure [dbar]

Residual = instrument oxygen - bath oxygen

Date, Delta Ox (ml/l)



**CALIBRATION
AFTER
MODIFICATIONS**