

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA
Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1329
CALIBRATION DATE: 19-Jun-08p

SBE 43 OXYGEN CALIBRATION DATA

COEFFICIENTS

Soc = 0.4348

Voffset = -0.4941

Tau20 = 1.57

A = -2.2326e-003

B = 1.8362e-004

C = -3.3714e-006

E nominal = 0.036

BATH OX (ml/l)	BATH TEMP ITS-90	BATH SAL PSU	INSTRUMENT OUTPUT(VOLTS)	INSTRUMENT OXYGEN(ml/l)	RESIDUAL (ml/l)
1.26	12.00	0.01	0.881	1.26	0.00
1.26	2.00	0.00	0.795	1.26	-0.00
1.27	6.00	0.00	0.831	1.27	0.00
1.27	20.00	0.01	0.953	1.27	-0.00
1.27	26.00	0.01	1.006	1.27	-0.00
1.29	30.00	0.01	1.051	1.29	-0.00
4.16	20.00	0.01	1.995	4.16	0.00
4.16	12.00	0.01	1.770	4.16	-0.00
4.17	26.00	0.01	2.171	4.17	0.00
4.17	6.00	0.00	1.604	4.17	0.00
4.17	2.00	0.00	1.489	4.17	-0.00
4.18	30.00	0.01	2.299	4.18	-0.00
6.52	30.00	0.01	3.310	6.52	-0.00
6.61	20.00	0.01	2.878	6.61	-0.01
6.62	26.00	0.01	3.162	6.63	0.01
6.65	12.00	0.01	2.535	6.65	0.00
6.67	2.00	0.00	2.085	6.67	-0.00
6.68	6.00	0.00	2.272	6.68	0.00

Oxygen (ml/l) = Soc * (V + Voffset) * (1.0 + A * T + B * T² + C * T³) * OxSol(T,S) * exp(E * P / K)

V = voltage output from SBE43, T = temperature [deg C], S = salinity [PSU] K = temperature [deg K]

OxSol(T,S) = oxygen saturation [ml/l], P = pressure [dbar], Residual = instrument oxygen - bath oxygen

Date, Delta Ox (ml/l)

