

# SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA  
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0140  
 CALIBRATION DATE: 15-Sep-09p

## SBE 43 OXYGEN CALIBRATION DATA

### COEFFICIENTS

Soc = 0.3185  
 Voffset = -0.6083  
 Tau20 = 1.56

A = -2.2375e-003  
 B = 1.4514e-004  
 C = -2.5553e-006  
 E nominal = 0.036


### NOMINAL DYNAMIC COEFFICIENTS

D1 = 1.92634e-4 H1 = -3.30000e-2  
 D2 = -4.64803e-2 H2 = 5.00000e+3  
 H3 = 1.45000e+3

BATH OX (ml/l)	BATH TEMP ITS-90	BATH SAL PSU	INSTRUMENT OUTPUT(VOLTS)	INSTRUMENT OXYGEN(ml/l)	RESIDUAL (ml/l)
1.23	2.00	0.00	1.010	1.23	-0.00
1.24	6.00	0.00	1.060	1.24	0.00
1.25	12.00	0.01	1.134	1.25	0.00
1.25	20.00	0.01	1.233	1.26	0.00
1.26	26.00	0.01	1.310	1.26	0.01
1.26	30.00	0.01	1.364	1.27	0.01
4.16	2.00	0.00	1.961	4.15	-0.01
4.16	26.00	0.01	2.924	4.17	0.00
4.17	30.00	0.01	3.098	4.17	0.00
4.17	20.00	0.01	2.680	4.17	-0.00
4.18	12.00	0.01	2.364	4.17	-0.01
4.18	6.00	0.00	2.126	4.17	-0.01
6.65	26.00	0.01	4.306	6.65	0.00
6.65	30.00	0.01	4.575	6.64	-0.00
6.71	20.00	0.01	3.939	6.70	-0.01
6.78	12.00	0.01	3.465	6.79	0.01
6.79	6.00	0.00	3.077	6.79	0.00
6.80	2.00	0.00	2.826	6.81	0.01

Oxygen (ml/l) = Soc \* (V + Voffset) \* (1.0 + A \* T + B \* T<sup>2</sup> + C \* T<sup>3</sup>) \* OxSol(T,S) \* exp(E \* P / K)  
 V = voltage output from SBE43, T = temperature [deg C], S = salinity [PSU] K = temperature [deg K]  
 OxSol(T,S) = oxygen saturation [ml/l], P = pressure [dbar], Residual = instrument oxygen - bath oxygen

Date, Delta Ox (ml/l)



15-Sep-09p 1.0000