



Sea-Bird Electronics, Inc.

1808 136th Place NE, Bellevue, Washington 98005 USA
 Website: <http://www.seabird.com>

Phone: (425) 643-9866
 FAX: (425) 643-9954
 Email: seabird@seabird.com

SBE Pressure Test Certificate

Test Date: 1/11/2000 Description SBE-3 Temperature Sensor

Job Number: 36257 Customer Name NOAA/AOML

SBE Sensor Information:

Model Number: 03
 Serial Number: 2958

Pressure Sensor Information:

Sensor Type: None
 Sensor Serial Number: None
 Sensor Rating: 0

Pressure Test Protocol:

Low Pressure Test: 50 PSI Held For 30 Minutes

High Pressure Test: 10000 PSI Held For 30 Minutes

Passed Test: ☒

Tested By: PC

