



SEA-BIRD ELECTRONICS, INC.

1808 - 136th Place Northeast, Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

Customer *US Department of Commerce, NOAA*

Job Number *46934*

Date *6/27/2007*

Technician *JMc*

Serial Number 042973

Low Pressure (PSI) 50 PSI

Time (Minutes) 15 Minutes

High Pressure (PSI) 10000 PSI

Time (Minutes) 45 Minutes

Pass ☒

Fail ☐

Comments

Replaced the main piston "O"-Rings and conductivity cell

