



Sea-Bird Electronics, Inc.

1808 136th Place NE, Bellevue, Washington 98005 USA

Website: <http://www.seabird.com>

Phone: (425) 643-9866

FAX: (425) 643-9954

Email: seabird@seabird.com

SBE Pressure Test Certificate

Test Date: 8/6/2007

Description SBE-43 DO Sensor

Job Number: 48081

Customer Name NOAA/AOML

SBE Sensor Information:

Model Number: 43

Serial Number: 1329

Pressure Sensor Information:

Sensor Type: None

Sensor Serial Number: None

Sensor Rating: 0

Pressure Test Protocol:

Low Pressure Test: 50 PSI Held For 15 Minutes

High Pressure Test: 10000 PSI Held For 30 Minutes

Passed Test: ☒

Tested By: mjd

