

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington 98005 USA
Phone: (425) 643 - 9866 Fax: (425) 643 - 9954 Internet: seabird@seabird.com

SENSOR SERIAL NUMBER = 1609
CALIBRATION DATE: 14-Nov-00s

TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

ITS-90 COEFFICIENTS

$g = 4.86579329e-03$
 $h = 6.79748362e-04$
 $i = 2.61064749e-05$
 $j = 2.00700808e-06$
 $f_0 = 1000.000$

IPTS-68 COEFFICIENTS

$a = 3.68126802e-03$
 $b = 6.03727652e-04$
 $c = 1.49587122e-05$
 $d = 2.00844439e-06$
 $f_0 = 6398.257$

BATH TEMP
(ITS-90 °C)

INSTRUMENT FREQ
(Hz)

INST TEMP
(ITS-90 °C)

RESIDUAL
(ITS-90 °C)

-1.5043	6398.257
1.0569	6773.832
4.6315	7324.164
8.1382	7894.408
11.6405	8494.712
15.2009	9137.249
18.6643	9794.053
22.1645	10490.684
25.6919	11226.700
29.1622	11984.687
32.6365	12777.800


-1.5041	0.00014
1.0568	-0.00012
4.6313	-0.00019
8.1381	-0.00005
11.6407	0.00016
15.2013	0.00043
18.6640	-0.00025
22.1643	-0.00019
25.6918	-0.00001
29.1623	0.00004
32.6365	0.00004

Temperature ITS-90 = $1/\{g + h[\ln(f_0/f)] + i[\ln^2(f_0/f)] + j[\ln^3(f_0/f)]\} - 273.15$ (°C)

Temperature IPTS-68 = $1/\{a + b[\ln(f_0/f)] + c[\ln^2(f_0/f)] + d[\ln^3(f_0/f)]\} - 273.15$ (°C)

Following the recommendation of JPOTS: T_{68} is assumed to be $1.00024 * T_{90}$ (-2 to 35 °C).

Residual = instrument temperature - bath temperature



POST CRUISE
CALIBRATION