

SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1329
CALIBRATION DATE: 26-Feb-11p

SBE 43 OXYGEN CALIBRATION DATA

COEFFICIENTS

Soc = 0.4191

Voffset = -0.4993

Tau20 = 1.33

A = -2.4144e-003

B = 1.0985e-004

C = -1.5952e-006

E nominal = 0.036

NOMINAL DYNAMIC COEFFICIENTS

D1 = 1.92634e-4 H1 = -3.30000e-2

D2 = -4.64803e-2 H2 = 5.00000e+3

H3 = 1.45000e+3


BATH OX (ml/l)	BATH TEMP ITS-90	BATH SAL PSU	INSTRUMENT OUTPUT(VOLTS)	INSTRUMENT OXYGEN(ml/l)	RESIDUAL (ml/l)
1.28	2.00	0.01	0.815	1.28	-0.00
1.28	6.00	0.01	0.854	1.28	0.00
1.28	26.00	0.02	1.048	1.28	-0.00
1.28	20.00	0.02	0.989	1.28	-0.00
1.29	12.00	0.01	0.913	1.29	0.00
1.29	30.00	0.02	1.093	1.29	0.00
4.19	6.00	0.01	1.659	4.19	-0.00
4.19	26.00	0.02	2.291	4.19	0.00
4.20	12.00	0.01	1.848	4.19	-0.00
4.20	20.00	0.02	2.100	4.20	-0.00
4.20	30.00	0.02	2.428	4.20	0.00
4.20	2.00	0.01	1.540	4.20	-0.00
6.64	30.00	0.02	3.542	6.63	-0.01
6.65	12.00	0.01	2.638	6.65	0.00
6.66	6.00	0.01	2.345	6.66	0.00
6.67	2.00	0.01	2.151	6.67	-0.00
6.68	20.00	0.02	3.048	6.68	-0.01
6.70	26.00	0.02	3.365	6.71	0.01

Oxygen (ml/l) = Soc * (V + Voffset) * (1.0 + A * T + B * T² + C * T³) * OxSol(T,S) * exp(E * P / K)

V = voltage output from SBE43, T = temperature [deg C], S = salinity [PSU] K = temperature [deg K]

OxSol(T,S) = oxygen saturation [ml/l], P = pressure [dbar], Residual = instrument oxygen - bath oxygen

Date, Delta Ox (ml/l)



26-Feb-11p 1.0000