

SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3338
CALIBRATION DATE: 15-Feb-11

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -1.01930182e+001
h = 1.57668781e+000
i = -2.56138918e-003
j = 2.98574515e-004
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 1.24221463e-006
b = 1.57004987e+000
c = -1.01797892e+001
d = -7.92357721e-005
m = 6.0
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.54631	0.00000	0.00000
-1.0000	34.9805	2.81657	4.94082	2.81657	0.00001
1.0000	34.9810	2.98870	5.05039	2.98869	-0.00001
15.0000	34.9802	4.28958	5.81165	4.28958	-0.00000
18.5000	34.9797	4.63770	5.99886	4.63771	0.00001
29.0000	34.9790	5.72596	6.54923	5.72594	-0.00002
32.5001	34.9723	6.10012	6.72794	6.10013	0.00001

Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

