



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

Customer US Department of Commerce, NOAA

Job Number 78180

Date 2/13/2014

Technician MD

Serial Number 431266

Low Pressure (PSI) 40 PSI

Time (Minutes) 15 Minutes

High Pressure (PSI) 10000 PSI

Time (Minutes) 30 Minutes

Pass ☒

Fail ☐

Comments

Replaced spar, getter, batteries and main piston O-rings.

