

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 4204
CALIBRATION DATE: 14-Oct-14

SBE 4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.70694888e+000
h = 1.46217411e+000
i = -1.86684661e-003
j = 2.34661074e-004

CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.57944	0.00000	0.00000
-1.0000	34.6727	2.79408	5.08018	2.79408	-0.00001
1.0000	34.6735	2.96492	5.19407	2.96493	0.00001
15.0001	34.6745	4.25606	5.98462	4.25606	-0.00000
18.5000	34.6747	4.60161	6.17890	4.60160	-0.00001
29.0001	34.6731	5.68151	6.74971	5.68153	0.00002
32.5000	34.6671	6.05291	6.93500	6.05289	-0.00002

f = INST FREQ / 1000.0

Conductivity = $(g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$ Siemens / meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = instrument conductivity - bath conductivity

