



Sea-Bird Scientific  
13431 NE 20<sup>th</sup> Street  
Bellevue, WA 98005  
USA

+1 425-643-9866  
seabird@seabird.com  
www.seabird.com

SENSOR SERIAL NUMBER: 0316  
CALIBRATION DATE: 18-Jan-19

Glider APL TEMPERATURE CALIBRATION DATA  
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

g = 4.43510456e-003  
h = 6.58756927e-004  
i = 2.72403037e-005  
j = 3.40132143e-006  
f0 = 1000.0

| BATH TEMP<br>(° C) | INSTRUMENT<br>OUTPUT (Hz) | INST TEMP<br>(° C) | RESIDUAL<br>(° C) |
|--------------------|---------------------------|--------------------|-------------------|
| 1.0000             | 3489.996                  | 1.0000             | -0.00002          |
| 4.5000             | 3765.342                  | 4.5000             | 0.00004           |
| 15.0000            | 4683.199                  | 15.0000            | 0.00001           |
| 18.5000            | 5020.869                  | 18.4999            | -0.00007          |
| 24.0000            | 5584.710                  | 24.0000            | 0.00001           |
| 29.0000            | 6133.282                  | 29.0001            | 0.00008           |
| 32.5000            | 6538.063                  | 32.5000            | -0.00005          |

f = Instrument Output (Hz)

Temperature ITS-90 (°C) =  $1/\{g + h[\ln(f_0 / f)] + i[\ln^2(f_0 / f)] + j[\ln^3(f_0 / f)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

