



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 0350
CALIBRATION DATE: 27-Jan-21

Glider APL CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.89711044e+000
h = 1.13378769e+000
i = -3.12285844e-003
j = 3.04620011e-004

CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2.96314	0.00000	0.00000
0.9999	34.6630	2.96410	5.92576	2.96411	0.00001
4.4999	34.6434	3.27001	6.15040	3.27000	-0.00000
14.9999	34.6027	4.24816	6.81851	4.24814	-0.00002
18.4999	34.5943	4.59208	7.03808	4.59206	-0.00002
23.9999	34.5855	5.14806	7.37896	5.14810	0.00003
29.0000	34.5815	5.66817	7.68377	5.66819	0.00002
32.5000	34.5799	6.03941	7.89391	6.03939	-0.00002

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

