

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3860
CALIBRATION DATE: 18-Mar-15

SBE 4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.03279941e+001
h = 1.48510533e+000
i = -1.17581430e-003
j = 1.67579701e-004

CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.63884	0.00000	0.00000
-1.0000	34.6628	2.79336	5.07860	2.79336	-0.00000
1.0000	34.6634	2.96414	5.19061	2.96414	0.00000
15.0000	34.6639	4.25489	5.96920	4.25489	-0.00000
18.5000	34.6637	4.60031	6.16077	4.60030	-0.00001
29.0000	34.6614	5.67980	6.72408	5.67982	0.00002
32.5000	34.6524	6.05063	6.90687	6.05062	-0.00001

f = INST FREQ / 1000.0

Conductivity = $(g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$ Siemens / meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = instrument conductivity - bath conductivity

