

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 5233
CALIBRATION DATE: 18-Jun-13

SBE3 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

ITS-90 COEFFICIENTS

g = 4.40263086e-003
h = 6.80263725e-004
i = 2.84109039e-005
j = 2.11443477e-006
f0 = 1000.0

IPTS-68 COEFFICIENTS

a = 3.68121193e-003
b = 6.25265181e-004
c = 2.14163250e-005
d = 2.11643619e-006
f0 = 3026.816

BATH TEMP (ITS-90)	INSTRUMENT FREQ (Hz)	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
-1.5000	3026.816	-1.5000	0.00003
1.0000	3194.113	1.0000	-0.00004
4.5000	3439.479	4.5000	-0.00002
8.0000	3698.177	8.0000	0.00001
11.5000	3970.565	11.5000	0.00002
15.0000	4256.994	15.0000	0.00002
18.5000	4557.805	18.5000	-0.00002
22.0000	4873.340	22.0000	-0.00001
25.5000	5203.915	25.5000	-0.00004
28.9999	5549.850	28.9999	0.00004
32.5000	5911.456	32.5000	-0.00001

Temperature ITS-90 = $1/\{g + h[\ln(f_0/f)] + i[\ln^2(f_0/f)] + j[\ln^3(f_0/f)]\} - 273.15$ (°C)

Temperature IPTS-68 = $1/\{a + b[\ln(f_0/f)] + c[\ln^2(f_0/f)] + d[\ln^3(f_0/f)]\} - 273.15$ (°C)

Following the recommendation of JPOTS: T_{68} is assumed to be $1.00024 * T_{90}$ (-2 to 35 °C)

Residual = instrument temperature - bath temperature

