



# SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 [www.seabird.com](http://www.seabird.com)

## Pressure Test Certificate

Customer US Department of Commerce, NOAA

Job Number 61048

Date 8/30/2010

Technician RH

Serial Number 430154

Low Pressure (PSI) 50 PSI

Time (Minutes) 15 Minutes

High Pressure (PSI) 10000 PSI

Time (Minutes) 30 Minutes

Pass ☒

Fail ☐

Comments

Replaced the spar and the main piston O-rings.

