

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3861
CALIBRATION DATE: 15-Aug-12

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Seimens/meter

GHIJ COEFFICIENTS

g = -1.02390722e+001
h = 1.36104375e+000
i = -6.41445745e-004
j = 1.17030313e-004
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 1.64664234e-005
b = 1.35958276e+000
c = -1.02364152e+001
d = -8.50547976e-005
m = 4.6
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.74369	0.00000	0.00000
-1.0000	34.7838	2.80220	5.30223	2.80219	-0.00002
1.0001	34.7841	2.97348	5.41951	2.97350	0.00002
15.0000	34.7857	4.26826	6.23449	4.26825	-0.00000
18.5000	34.7855	4.61472	6.43498	4.61473	0.00000
29.0001	34.7841	5.69765	7.02447	5.69764	-0.00000
32.5001	34.7767	6.06988	7.21587	6.06988	0.00000


Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction



25-Feb-11 0.9999898
15-Aug-12 1.0000000