

Conductivity Calibration Report

Customer:	US Department of Commerce, NOAA				
Job Number:	70501	Date of Report:	8/15/2012		
Model Number	SBE 04C	Serial Number:	042973		

Conductivity sensors are normally calibrated 'as received', without cleaning or adjustments, allowing a determination of sensor drift. If the calibration identifies a problem or indicates cell cleaning is necessary, then a second calibration is performed after work is completed. The 'as received' calibration is not performed if the sensor is damaged or non-functional, or by customer request.

An 'as received' calibration certificate is provided, listing the coefficients used to convert sensor frequency to conductivity. Users must choose whether the 'as received' calibration or the previous calibration better represents the sensor condition during deployment. In SEASOFT enter the chosen coefficients. The coefficient 'slope' allows small corrections for drift between calibrations (consult the SEASOFT manual). Calibration coefficients obtained after a repair or cleaning apply only to subsequent data.

'AS RECEIVED CALIBRATION'	✓ Per	formed	□ Not Performed	
Date: 8/15/2012	Drift since last cal:	+0.0	00000] PSU/month*
Comments:				

'CALIBRA'	TION AFT	'ER CLEANING & REPLATINIZING'	Perf	ormed	Not	Performed
Date:		Drift since La	ast cal:			PSU/month*

Comments:

*Measured at 3.0 S/m

Cell cleaning and electrode replatinizing tend to 'reset' the conductivity sensor to its original condition. Lack of drift in post-cleaning-calibration indicates geometric stability of the cell and electrical stability of the sensor circuit.