

## Sea-Bird Electronics, Inc.

13431 NE 20th St. Bellevue, Washington 98005 USA Website: http://www.seabird.com

Phone: (425) 643-9866 FAX: (425) 643-9954 Email: seabird@seabird.com

## **SBE Pressure Test Certificate**

Test Date: <u>10/9/2013</u> Description <u>SBE-5T Submersible Pump</u>

Job Number: **STOCK** Customer Name

**SBE Sensor Information:** Pressure Sensor Information:

Model Number: <u>5T</u> Sensor Type: <u>None</u>

Serial Number: 7274 Sensor Serial Number: None

Sensor Rating:  $\underline{\mathbf{0}}$ 

## **Pressure Test Protocol:**

Low Pressure Test: 40 PSI Held For 15 Minutes

High Pressure Test: 10000 PSI Held For 15 Minutes

Passed Test:

Tested By: ND

