

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 2082
CALIBRATION DATE: 15-Dec-15

SBE 43 OXYGEN CALIBRATION DATA

COEFFICIENTS:

Soc = 0.4207
Voffset = -0.5332
Tau20 = 1.05

A = -3.3688e-003
B = 2.1886e-004
C = -3.2294e-006
E nominal = 0.036

NOMINAL DYNAMIC COEFFICIENTS

D1 = 1.92634e-4 H1 = -3.300000e-2
D2 = -4.64803e-2 H2 = 5.00000e+3
H3 = 1.45000e+3

BATH OXYGEN (ml/l)	BATH TEMPERATURE (° C)	BATH SALINITY (PSU)	INSTRUMENT OUTPUT (volts)	INSTRUMENT OXYGEN (ml/l)	RESIDUAL (ml/l)
1.15	6.00	0.00	0.851	1.15	-0.00
1.16	2.00	0.00	0.819	1.16	-0.00
1.17	12.00	0.00	0.906	1.17	-0.00
1.17	26.00	0.00	1.023	1.17	0.00
1.17	30.00	0.00	1.056	1.17	0.00
1.17	20.00	0.00	0.975	1.18	0.00
3.91	26.00	0.00	2.164	3.91	0.00
3.91	20.00	0.00	2.005	3.92	0.00
3.92	30.00	0.00	2.281	3.92	0.00
3.92	6.00	0.00	1.618	3.92	-0.00
3.93	2.00	0.00	1.504	3.93	0.00
3.94	12.00	0.00	1.792	3.94	0.00
6.66	6.00	0.00	2.374	6.66	-0.00
6.68	12.00	0.00	2.671	6.68	-0.00
6.70	2.00	0.00	2.188	6.70	0.00
6.70	30.00	0.00	3.520	6.71	0.00
6.72	20.00	0.00	3.061	6.73	0.00
6.76	26.00	0.00	3.351	6.75	-0.01

V = instrument output (volts); T = temperature (°C); S = salinity (PSU); K = temperature (°K)

Oxsol(T,S) = oxygen saturation (ml/l); P = pressure (dbar)

Oxygen (ml/l) = Soc * (V + Voffset) * (1.0 + A * T + B * T² + C * T³) * Oxsol(T,S) * exp(E * P / K)

Residual (ml/l) = instrument oxygen - bath oxygen

