



Sea-Bird Scientific  
13431 NE 20<sup>th</sup> Street  
Bellevue, WA 98005  
USA

+1 425-643-9866  
seabird@seabird.com  
www.seabird.com

SENSOR SERIAL NUMBER: 0317  
CALIBRATION DATE: 18-Feb-21

Glider APL TEMPERATURE CALIBRATION DATA  
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

g = 4.40252177e-003  
h = 6.40847403e-004  
i = 2.64784058e-005  
j = 3.35854965e-006  
f0 = 1000.0

| BATH TEMP<br>(° C) | INSTRUMENT<br>OUTPUT (Hz) | INST TEMP<br>(° C) | RESIDUAL<br>(° C) |
|--------------------|---------------------------|--------------------|-------------------|
| 1.0000             | 3423.834                  | 0.9999             | -0.00007          |
| 4.5000             | 3701.430                  | 4.5001             | 0.00013           |
| 15.0000            | 4630.574                  | 14.9999            | -0.00008          |
| 18.4999            | 4973.717                  | 18.4998            | -0.00005          |
| 24.0000            | 5548.082                  | 24.0000            | 0.00003           |
| 29.0000            | 6108.394                  | 29.0001            | 0.00015           |
| 32.5001            | 6522.697                  | 32.5000            | -0.00010          |

f = Instrument Output (Hz)

Temperature ITS-90 (°C) =  $1 / \{g + h[\ln(f_0 / f)] + i[\ln^2(f_0 / f)] + j[\ln^3(f_0 / f)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

