



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 0351
CALIBRATION DATE: 09-Feb-22

Glider APL CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.00284659e+001
h = 1.13373497e+000
i = -3.63966628e-003
j = 3.32101131e-004

CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2.98456	0.00000	0.00000
1.0000	34.5582	2.95600	5.93521	2.95604	0.00004
4.4999	34.5396	3.26117	6.15947	3.26113	-0.00004
15.0000	34.4994	4.23683	6.82659	4.23675	-0.00008
18.5000	34.4876	4.57944	7.04572	4.57953	0.00009
24.0000	34.4783	5.13387	7.38605	5.13388	0.00001
29.0000	34.4728	5.65235	7.69036	5.65234	-0.00002
32.5001	34.4677	6.02205	7.89999	6.02204	-0.00001

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

