



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

Customer US Department of Commerce, NOAA

Job Number 84659

Date 5/20/2015

Technician BLT

Serial Number 3236257-0500

Low Pressure (PSI) 50 PSI

Time (Minutes) 60 Minutes

High Pressure (PSI) *N/A PSI

Time (Minutes) *N/A Minutes

Pass ☒

Fail ☐

Comments

Replaced the main piston "O"-Rings. Replaced 6-pin connector.

* Due to the dimensions of the CAROUSEL we are unable to perform a high pressure test at this time.

