



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-0886
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 1387
CALIBRATION DATE: 11-May-17

SBE 4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.06544844e+001
h = 1.59467817e+000
i = -4.85925254e-004
j = 1.23503810e-004

CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
0.0000	0.0000	0.00000	2.58517	0.00000	0.00000
-1.0000	34.6347	2.79131	4.91691	2.79129	-0.00002
1.0000	34.6355	2.96198	5.02443	2.96199	0.00001
15.0000	34.6352	4.25174	5.77212	4.25176	0.00001
18.5000	34.6339	4.59678	5.95614	4.59677	-0.00001
29.0000	34.6256	5.67459	6.49728	5.67457	-0.00002
32.5000	34.6103	6.04412	6.67264	6.04413	0.00001

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

