



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 1387
CALIBRATION DATE: 12-Mar-21

SBE 4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.06488655e+001
h = 1.59113254e+000
i = 2.71219676e-004
j = 9.32461560e-005

CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
0.0000	0.0000	0.00000	2.58594	0.00000	0.00000
-1.0001	34.7258	2.79796	4.92163	2.79793	-0.00002
0.9999	34.7250	2.96890	5.02919	2.96892	0.00003
15.0001	34.7228	4.26137	5.77734	4.26135	-0.00002
18.4998	34.7195	4.60689	5.96136	4.60690	0.00001
29.0000	34.7086	5.68666	6.50262	5.68666	0.00000
32.4998	34.6887	6.05623	6.67770	6.05623	-0.00000

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

