



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 1387
CALIBRATION DATE: 27-Apr-22

SBE 4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.06597064e+001
h = 1.59695695e+000
i = -1.23501679e-003
j = 1.88594469e-004

CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
0.0000	0.0000	0.00000	2.58517	0.00000	0.00000
-1.0001	34.6626	2.79334	4.91828	2.79333	-0.00001
0.9999	34.6624	2.96405	5.02579	2.96406	0.00001
15.0000	34.6606	4.25453	5.77368	4.25453	-0.00000
18.4999	34.6578	4.59960	5.95766	4.59960	0.00000
29.0000	34.6493	5.67804	6.49887	5.67803	-0.00000
32.4999	34.6347	6.04788	6.67427	6.04789	0.00000

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

