



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 3854
CALIBRATION DATE: 13-Dec-18

SBE 4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.04239032e+001
h = 1.58565270e+000
i = -2.07124491e-003
j = 2.54727419e-004

CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
0.0000	0.0000	0.00000	2.56691	0.00000	0.00000
-1.0000	34.5297	2.78363	4.91836	2.78362	-0.00001
1.0000	34.5298	2.95380	5.02649	2.95381	0.00001
15.0000	34.5285	4.24003	5.77824	4.24003	0.00000
18.5000	34.5269	4.58410	5.96316	4.58409	-0.00001
29.0000	34.5191	5.65909	6.50681	5.65911	0.00002
32.5000	34.5075	6.02820	6.68315	6.02819	-0.00001

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

