



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 3854
CALIBRATION DATE: 23-Jun-21

SBE 4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.04011185e+001
h = 1.57823375e+000
i = -2.33538047e-004
j = 1.39385519e-004

CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
0.0000	0.0000	0.00000	2.56691	0.00000	0.00000
-1.0000	34.8209	2.80491	4.93237	2.80491	0.00000
1.0000	34.8212	2.97634	5.04094	2.97634	-0.00000
15.0000	34.8212	4.27215	5.79569	4.27215	-0.00000
18.5000	34.8199	4.61879	5.98133	4.61880	0.00000
29.0001	34.8130	5.70185	6.52707	5.70185	-0.00000
32.5001	34.7992	6.07336	6.70392	6.07336	0.00000

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

