



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 3861
CALIBRATION DATE: 11-Oct-17

SBE 4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.02522564e+001
h = 1.36416891e+000
i = -1.39690340e-003
j = 1.69935120e-004

CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
0.0000	0.0000	0.00000	2.74399	0.00000	0.00000
-1.0001	34.6803	2.79463	5.29673	2.79462	-0.00001
0.9999	34.6803	2.96544	5.41378	2.96545	0.00001
14.9999	34.6803	4.25668	6.22738	4.25667	-0.00001
18.4999	34.6790	4.60211	6.42747	4.60212	0.00001
28.9999	34.6719	5.68131	7.01554	5.68131	-0.00000
32.4999	34.6577	6.05144	7.20606	6.05145	0.00000

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

