



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

Pressure Test Certificate

Test Date: **2018-01-09**

Description: **SBE-21 Thermosalinograph**

Sensor Information:

Replaced the main piston "O"-Rings.

Model Number: **SBE-21**

Serial Number: **3269**

Pressure Test Protocol:

Low Pressure Test: **30** PSI Held For: **30** Minutes

High Pressure Test: **0** PSI Held For: **0** Minutes

Passed Test: **True**

Tested By: **TN**

