


Calibration Record

Date: February 6, 2018
Sensor Type: Seapoint Chlorophyll Fluorometer
Serial Number(s): 2985

A comparative calibration was performed using a calibrated reference fluorometer. The reference fluorometer was calibrated with the cultured algae *Isochrysis galbana*.

This instrument passes or exceeds the specifications stated in the supplied User Manual.

If you have any questions please contact me at 603/642-4921 or seapoint@seapoint.com

	02/06/18
Signature	Date



PO Box 368 • Exeter, NH 03833 • USA
Tel: (603) 642-4921 • Fax: (603) 642-4922
seapoint@seapoint.com • www.seapoint.com