



# SEA-BIRD ELECTRONICS, INC.

1808 - 136th Place Northeast, Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

## Pressure Test Certificate

Customer US Department of Commerce, NOAA

Job Number 39917

Date 8/9/2005

Technician BLT

Serial Number 041347

Low Pressure (PSI) 50 PSI

Time (Minutes) 15 Minutes

High Pressure (PSI) 10000 PSI

Time (Minutes) 30 Minutes

Pass ☒

Fail ☐

Comments

Replaced the conductivity cell and the main piston "O"-Rings.

