



Sea-Bird Electronics, Inc.

1808 136th Place NE, Bellevue, Washington 98005 USA

Website: <http://www.seabird.com>

Phone: (425) 643-9866

FAX: (425) 643-9954

Email: seabird@seabird.com

SBE Pressure Test Certificate

Test Date: 9/15/2009

Description SBE-9Plus CTD

Job Number: 54833

Customer Name NOAA/AOML

SBE Sensor Information:

Model Number: 09

Serial Number: 0957

Pressure Sensor Information:

Sensor Type: DigiQuartz

Sensor Serial Number: 115173

Sensor Rating: 10000

Pressure Test Protocol:

Low Pressure Test: 40 PSI Held For 15 Minutes

High Pressure Test: 10000 PSI Held For 15 Minutes

Passed Test: ☒

Tested By: RL

