

Pressure Test Certificate

Customer US Department of Commerce, NOAA
Job Number 39917
Date 8/9/2005
Technician BLT

Serial Number 041335

Low Pressure (PSI) 50 PSI
Time (Minutes) 15 Minutes

High Pressure (PSI) 10000 PSI
Time (Minutes) 30 Minutes

Pass ☒
Fail ☐

Comments
Replaced the conductivity cell and the main piston "O"-Rings.

