

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3657
CALIBRATION DATE: 14-Aug-09

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Seimens/meter

GHIJ COEFFICIENTS

g = -9.89444754e+000
h = 1.40009198e+000
i = -2.48419181e-003
j = 2.56836805e-004
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 2.88648139e-007
b = 1.39305191e+000
c = -9.87828742e+000
d = -7.66918769e-005
m = 6.5
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.66295	0.00000	0.00000
-1.0000	34.6992	2.79602	5.21085	2.79601	-0.00001
1.0000	34.6994	2.96692	5.32715	2.96693	0.00001
15.0000	34.7011	4.25897	6.13494	4.25897	-0.00000
18.5000	34.7010	4.60472	6.33352	4.60474	0.00002
29.0000	34.7001	5.68542	6.91703	5.68539	-0.00003
32.5000	34.6949	6.05721	7.10659	6.05723	0.00002


Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction



14-Aug-09 1.0000000