



Sea-Bird Electronics, Inc.
1808 136th Place NE, Bellevue, Washington 98005 USA
Website: <http://www.seabird.com>

Phone: (425) 643-9866
FAX: (425) 643-9954
Email: seabird@seabird.com

SBE Pressure Test Certificate

Test Date: 5/17/2004

Description SBE-9Plus CTD

Job Number: 36257

Customer Name NOAA/AOML

SBE Sensor Information:
Model Number: 09
Serial Number: 0768

Pressure Sensor Information:
Sensor Type: DigiQuartz
Sensor Serial Number: 92973
Sensor Rating: 10000

Pressure Test Protocol:

Low Pressure Test: 50 PSI Held For 15 Minutes

High Pressure Test: 10000 PSI Held For 15 Minutes

Passed Test: ☒

Tested By: PC

