

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington 98005 USA
Phone: (425) 643 - 9866 Fax: (425) 643 - 9954 Internet: seabird@seabird.com

SENSOR SERIAL NUMBER = 1075
CALIBRATION DATE: 16-Nov-99s

TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

ITS-90 COEFFICIENTS

$g = 4.81091253e-03$
 $h = 6.68979182e-04$
 $i = 2.51679163e-05$
 $j = 1.99205337e-06$
 $f_0 = 1000.000$

IPTS-68 COEFFICIENTS

$a = 3.68139015e-03$
 $b = 5.98109917e-04$
 $c = 1.44847504e-05$
 $d = 1.99343776e-06$
 $f_0 = 6002.186$


BATH TEMP (ITS-90 °C)	INSTRUMENT FREQ (Hz)	INST TEMP (ITS-90 °C)	RESIDUAL (ITS-90 °C)
-1.5131	6002.186	-1.5131	0.00000
1.0483	6357.997	1.0483	-0.00000
4.6212	6879.371	4.6212	0.00001
8.1280	7420.158	8.1280	0.00001
11.6314	7989.978	11.6314	-0.00004
15.1916	8600.031	15.1915	-0.00002
18.6552	9224.262	18.6552	0.00004
22.1560	9886.639	22.1560	0.00004
25.6834	10586.721	25.6834	-0.00002
29.1546	11308.276	29.1546	-0.00006
32.6298	12063.713	32.6298	0.00003

Temperature ITS-90 = $1/\{g + h[\ln(f_0/f)] + i[\ln^2(f_0/f)] + j[\ln^3(f_0/f)]\} - 273.15$ (°C)

Temperature IPTS-68 = $1/\{a + b[\ln(f_0/f)] + c[\ln^2(f_0/f)] + d[\ln^3(f_0/f)]\} - 273.15$ (°C)

Following the recommendation of JPOTS: T_{68} is assumed to be $1.00024 * T_{90}$ (-2 to 35 °C).

Residual = instrument temperature - bath temperature



calibration date	delta T (mdeg C)
⊕ 22-Apr-99s	0.94
○ 16-Nov-99s	0.00

PORT CRUISE
CALIBRATION