

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 5140
CALIBRATION DATE: 10-Sep-09

SBE3 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

ITS-90 COEFFICIENTS

g = 4.36465389e-003
h = 6.40962773e-004
i = 2.22912766e-005
j = 2.07846363e-006
f0 = 1000.0

IPTS-68 COEFFICIENTS

a = 3.68121197e-003
b = 5.99478759e-004
c = 1.54325511e-005
d = 2.07993634e-006
f0 = 3017.194

BATH TEMP (ITS-90)	INSTRUMENT FREQ (Hz)	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
-1.5000	3017.194	-1.5000	0.00003
1.0000	3191.265	1.0000	-0.00003
4.5000	3447.006	4.5000	-0.00001
8.0000	3717.146	8.0000	-0.00002
11.5000	4002.079	11.5000	0.00003
15.0000	4302.173	15.0000	0.00004
18.5000	4617.790	18.5000	-0.00001
22.0000	4949.294	22.0000	-0.00001
25.5000	5297.021	25.5000	-0.00001
29.0001	5661.305	29.0001	-0.00003
32.5000	6042.438	32.5000	0.00002

Temperature ITS-90 = $1/[g + h[\ln(f_0/f)] + i[\ln^2(f_0/f)] + j[\ln^3(f_0/f)]] - 273.15$ (°C)

Temperature IPTS-68 = $1/[a + b[\ln(f_0/f)] + c[\ln^2(f_0/f)] + d[\ln^3(f_0/f)]] - 273.15$ (°C)

Following the recommendation of JPOTS: T_{68} is assumed to be $1.00024 * T_{90}$ (-2 to 35 °C)

Residual = instrument temperature - bath temperature

