



Sea-Bird Electronics, Inc.

1808 136th Place NE, Bellevue, Washington 98005 USA  
Website: <http://www.seabird.com>

Phone: (425) 643-9866  
FAX: (425) 643-9954  
Email: [seabird@seabird.com](mailto:seabird@seabird.com)

## SBE Pressure Test Certificate

Test Date: 10/10/2005 Description SBE-3 Temperature Sensor

Job Number: 47193 Customer Name NOAA/AOML

### SBE Sensor Information:

Model Number: 03  
Serial Number: 4799

### Pressure Sensor Information:

Sensor Type: None  
Sensor Serial Number: None  
Sensor Rating: 0

### Pressure Test Protocol:

Low Pressure Test: 40 PSI Held For 15 Minutes

High Pressure Test: 10000 PSI Held For 15 Minutes

Passed Test: ☒

Tested By: DF

