

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1692
CALIBRATION DATE: 13-Mar-09

SBE3 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

ITS-90 COEFFICIENTS

g = 4.80212798e-003
h = 6.72222918e-004
i = 2.57232804e-005
j = 2.04053148e-006
f0 = 1000.0

IPTS-68 COEFFICIENTS

a = 3.68121378e-003
b = 6.00468761e-004
c = 1.49115488e-005
d = 2.04195946e-006
f0 = 5874.353

BATH TEMP (ITS-90)	INSTRUMENT FREQ (Hz)	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
-1.5001	5874.353	-1.5001	-0.00001
0.9999	6212.676	0.9999	-0.00001
4.4999	6709.605	4.4999	0.00003
7.9999	7234.359	7.9999	0.00002
11.4999	7787.675	11.4999	0.00000
14.9999	8370.277	14.9999	0.00004
18.4999	8982.803	18.4997	-0.00018
21.9999	9626.039	21.9999	0.00004
25.4999	10300.523	25.5000	0.00009
28.9999	11006.873	28.9999	0.00003
32.4999	11745.696	32.4998	-0.00005

Temperature ITS-90 = $1/\{g + h[\ln(f_0/f)] + i[\ln^2(f_0/f)] + j[\ln^3(f_0/f)]\} - 273.15$ (°C)

Temperature IPTS-68 = $1/\{a + b[\ln(f_0/f)] + c[\ln^2(f_0/f)] + d[\ln^3(f_0/f)]\} - 273.15$ (°C)

Following the recommendation of JPOTS: T_{68} is assumed to be $1.00024 * T_{90}$ (-2 to 35 °C)

Residual = instrument temperature - bath temperature

