

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 2958
CALIBRATION DATE: 14-Aug-07

SBE3 TEMPERATURE CALIBRATION DATA
IPTS-90 TEMPERATURE SCALE

ITS-90 COEFFICIENTS

g = 4.39475142e-003
h = 6.72710675e-004
i = 2.98713927e-005
j = 2.75340725e-006
f0 = 1000.0

ITS-68 COEFFICIENTS

a = 3.68121080e-003
b = 6.16731556e-004
c = 2.07340280e-005
d = 2.75542016e-006
f0 = 3033.746

BATH TEMP (ITS-90)	INSTRUMENT FREQ (Hz)	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
-1.5000	3033.746	-1.4999	0.00011
1.0000	3203.796	0.9999	-0.00012
4.5000	3453.427	4.4999	-0.00007
8.0000	3716.859	8.0000	-0.00005
11.5000	3994.477	11.5001	0.00010
15.0000	4286.625	15.0001	0.00013
18.5000	4593.653	18.5000	0.00003
22.0000	4915.913	21.9999	-0.00007
25.5000	5253.741	25.4999	-0.00010
29.0000	5607.458	29.0000	-0.00003
32.5000	5977.363	32.5001	0.00008

Temperature ITS-90 = $1/\{g + h[\ln(f_0/f)] + i[\ln^2(f_0/f)] + j[\ln^3(f_0/f)]\} - 273.15$ (°C)

Temperature IPTS-68 = $1/\{a + b[\ln(f_0/f)] + c[\ln^2(f_0/f)] + d[\ln^3(f_0/f)]\} - 273.15$ (°C)

Following the recommendation of JPOTS: T_{68} is assumed to be $1.00024 * T_{90}$ (-2 to 35 °C)

Residual = instrument temperature - bath temperature

