

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1374
CALIBRATION DATE: 27-Feb-08

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -3.96378799e+000
h = 4.83835005e-001
i = -1.00960283e-004
j = 3.21292999e-005
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 1.53072754e-005
b = 4.83597262e-001
c = -3.96348723e+000
d = -9.06916192e-005
m = 4.2
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.86232	0.00000	0.00000
-1.0000	34.9366	2.81336	8.13394	2.81336	-0.00000
1.0000	34.9367	2.98527	8.34816	2.98528	0.00001
15.0000	34.9377	4.28493	9.81523	4.28489	-0.00003
18.5000	34.9373	4.63268	10.17138	4.63272	0.00003
29.0000	34.9359	5.71970	11.21002	5.71969	-0.00000
32.5000	34.9302	6.09360	11.54516	6.09360	-0.00000

Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

