

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0140
 CALIBRATION DATE: 19-Aug-09p

SBE 43 OXYGEN CALIBRATION DATA

COEFFICIENTS

Soc = 0.3091

Voffset = -0.6163

Tau20 = 1.11

A = -3.0551e-003

B = 2.9938e-004

C = -5.9407e-006

E nominal = 0.036

NOMINAL DYNAMIC COEFFICIENTS

D1 = 1.92634e-4 H1 = -3.30000e-2

D2 = -4.64803e-2 H2 = 5.00000e+3

H3 = 1.45000e+3

BATH OX (ml/l)	BATH TEMP ITS-90	BATH SAL PSU	INSTRUMENT OUTPUT(VOLTS)	INSTRUMENT OXYGEN(ml/l)	RESIDUAL (ml/l)
1.25	2.00	0.00	1.036	1.25	-0.00
1.25	6.00	0.00	1.086	1.25	-0.00
1.26	12.00	0.01	1.160	1.26	0.00
1.27	20.00	0.01	1.255	1.27	0.00
1.28	26.00	0.01	1.332	1.28	0.00
1.28	30.00	0.01	1.386	1.28	0.00
4.20	20.00	0.01	2.729	4.20	0.00
4.20	30.00	0.01	3.142	4.20	0.00
4.20	2.00	0.00	2.028	4.20	-0.00
4.20	26.00	0.01	2.966	4.20	-0.00
4.21	6.00	0.00	2.190	4.20	-0.01
4.21	12.00	0.01	2.429	4.21	0.00
6.60	30.00	0.01	4.584	6.60	0.00
6.64	26.00	0.01	4.327	6.63	-0.00
6.66	20.00	0.01	3.966	6.66	0.00
6.73	12.00	0.01	3.517	6.73	0.00
6.76	6.00	0.00	3.151	6.76	-0.00
6.80	2.00	0.00	2.901	6.80	0.00

Oxygen (ml/l) = Soc * (V + Voffset) * (1.0 + A * T + B * T² + C * T³) * OxSol(T,S) * exp(E * P / K)

V = voltage output from SBE43, T = temperature [deg C], S = salinity [PSU] K = temperature [deg K]

OxSol(T,S) = oxygen saturation [ml/l], P = pressure [dbar], Residual = instrument oxygen - bath oxygen

Date, Delta Ox (ml/l)

