

# SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1374  
CALIBRATION DATE: 21-May-10

SBE4 CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Seimens/meter

## GHIJ COEFFICIENTS

g = -3.43528594e+000  
h = 3.62314202e-001  
i = 2.32682484e-002  
j = -1.15825685e-003  
CPcor = -9.5700e-008 (nominal)  
CTcor = 3.2500e-006 (nominal)

## ABCDM COEFFICIENTS

a = -1.29669816e-043  
b = 4.86433181e-001  
c = -4.18472212e+000  
d = -2.86391263e-003  
m = 40.1  
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.86269	0.00020	0.00020
-1.0000	34.6320	2.79111	8.12099	2.78839	-0.00272
0.9999	34.6324	2.96173	8.33450	2.96146	-0.00027
15.0000	34.6325	4.25145	9.80012	4.25770	0.00625
18.4999	34.6319	4.59653	10.15712	4.59953	0.00300
29.0000	34.6304	5.67529	11.20994	5.65759	-0.01769
32.5000	34.6241	6.04625	11.59206	6.05749	0.01124

Conductivity =  $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$  Siemens/meter

Conductivity =  $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$  Siemens/meter

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

