

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0154
 CALIBRATION DATE: 11-Aug-07p

SBE 43 OXYGEN CALIBRATION DATA

COEFFICIENTS

Soc = 0.4268

Boc = 0.0000

Voffset = -0.5100

TCor = 0.0005

PCor = 1.350e-04

BATH OX (ml/l)	BATH TEMP ITS-90	BATH SAL PSU	INSTRUMENT OUTPUT(VOLTS)	INSTRUMENT OXYGEN(ml/l)	RESIDUAL (ml/l)
1.24	2.00	0.00	0.808	1.23	-0.00
1.24	6.00	0.00	0.844	1.24	0.00
1.25	12.00	0.01	0.897	1.25	0.00
1.25	20.00	0.01	0.968	1.25	0.00
1.26	26.00	0.01	1.022	1.25	-0.00
1.27	30.00	0.01	1.070	1.28	0.00
4.13	26.00	0.01	2.193	4.12	-0.01
4.14	12.00	0.01	1.794	4.15	0.01
4.15	6.00	0.00	1.625	4.15	0.00
4.15	20.00	0.01	2.025	4.15	-0.00
4.16	2.00	0.00	1.513	4.14	-0.01
4.18	30.00	0.01	2.343	4.19	0.01
6.56	26.00	0.01	3.183	6.54	-0.02
6.57	20.00	0.01	2.908	6.56	-0.01
6.59	12.00	0.01	2.554	6.61	0.02
6.60	6.00	0.00	2.287	6.62	0.02
6.62	2.00	0.00	2.109	6.60	-0.02
7.07	30.00	0.01	3.608	7.07	0.01

oxygen (ml/l) = (Soc * (V + Voffset)) * exp(Tcor * T) * Oxsat(T,S) * exp(PCor * P)

V = voltage output from SBE43, T = temperature [deg C], S = salinity [PSU]

Oxsat(T,S) = oxygen saturation [ml/l], P = pressure [dbar]

Residual = instrument oxygen - bath oxygen

